

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
 Data File : VX001786.D
 Acq On : 17 May 2018 13:19
 Operator : JC/MD
 Sample : J2853-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW336-180507

Quant Time: May 18 04:29:27 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	300627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	398091	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	331926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172229	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	211089	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	5.51	113	180191	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
50) Toluene-d8	8.72	98	628705	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	11.15	95	186746	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	1696	0.52	ug/l	89
27) cis-1,2-Dichloroethene	4.60	96	92368	24.80	ug/l	90
44) Trichloroethene	7.22	130	57553	13.73	ug/l	97
64) Tetrachloroethene	9.34	164	1409097	326.11	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
Data File : VX001786.D
Acq On : 17 May 2018 13:19
Operator : JC/MD
Sample : J2853-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW336-180507

Quant Time: May 18 04:29:27 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

