

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
 Data File : VX001788.D
 Acq On : 17 May 2018 14:13
 Operator : JC/MD
 Sample : J2853-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY063MW003-180508

Quant Time: May 18 04:33:21 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	291911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	396757	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	155194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	211089	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
35) Dibromofluoromethane	5.51	113	179386	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
50) Toluene-d8	8.72	98	623287	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.14	95	175825	40.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.44%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.15	96	1082	0.34	ug/l	# 79
27) cis-1,2-Dichloroethene	4.60	96	320821	88.70	ug/l	90
44) Trichloroethene	7.22	130	434232	103.96	ug/l	97
64) Tetrachloroethene	9.34	164	832893	202.79	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
Data File : VX001788.D
Acq On : 17 May 2018 14:13
Operator : JC/MD
Sample : J2853-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY063MW003-180508

Quant Time: May 18 04:33:21 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

