

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
 Data File : VX001789.D
 Acq On : 17 May 2018 14:39
 Operator : JC/MD
 Sample : J2853-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW026-180507

Quant Time: May 18 04:34:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	294003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	397444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	320087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	159836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207706	52.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.42%	
35) Dibromofluoromethane	5.51	113	177884	53.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.08%	
50) Toluene-d8	8.72	98	623182	51.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.76%	
62) 4-Bromofluorobenzene	11.14	95	178840	41.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.70%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	2554	0.85	ug/l	95
21) trans-1,2-Dichloroethene	3.16	96	791	0.25	ug/l	# 78
27) cis-1,2-Dichloroethene	4.60	96	66418	18.23	ug/l	92
30) Chloroform	5.21	83	1420	0.24	ug/l	97
44) Trichloroethene	7.22	130	17074	4.08	ug/l	93
64) Tetrachloroethene	9.34	164	272271	65.34	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
Data File : VX001789.D
Acq On : 17 May 2018 14:39
Operator : JC/MD
Sample : J2853-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW026-180507

Quant Time: May 18 04:34:44 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

