

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
 Data File : VX001792.D
 Acq On : 17 May 2018 15:59
 Operator : JC/MD
 Sample : J2853-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0455-180507

Quant Time: May 18 04:37:57 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	296536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	393495	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	314668	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	205786	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
35) Dibromofluoromethane	5.51	113	176361	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
50) Toluene-d8	8.72	98	614613	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
62) 4-Bromofluorobenzene	11.14	95	179003	41.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.60%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	2673	0.88	ug/l	99
17) Carbon Disulfide	2.57	76	3111	0.41	ug/l #	87
21) trans-1,2-Dichloroethene	3.17	96	1383	0.43	ug/l #	84
27) cis-1,2-Dichloroethene	4.60	96	64940	17.68	ug/l	90
30) Chloroform	5.21	83	1524	0.25	ug/l	96
44) Trichloroethene	7.22	130	15444	3.73	ug/l	98
64) Tetrachloroethene	9.34	164	273729	66.82	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051718\
Data File : VX001792.D
Acq On : 17 May 2018 15:59
Operator : JC/MD
Sample : J2853-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0455-180507

Quant Time: May 18 04:37:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

