

Data Path : W:\HPCHEM1\MSVOA X\Data\VX051718\
 Data File : VX001803.D
 Acq On : 17 May 2018 20:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:59:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.502	0.460	8.4	84	0.00
3 P	Chloromethane	0.542	0.532	1.8	95	0.00
4 C	Vinyl Chloride	0.586	0.573	2.2#	92	0.00
5 T	Bromomethane	0.295	0.220	25.4#	81	0.00
6 T	Chloroethane	0.369	0.372	-0.8	92	0.00
7 T	Trichlorofluoromethane	0.852	0.803	5.8	88	0.00
8 T	Diethyl Ether	0.351	0.346	1.4	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.513	0.487	5.1	91	0.00
10 T	Methyl Iodide	0.472	0.651	-37.9#	122	0.00
11 T	Tert butyl alcohol	0.119	0.132	-10.9	106	0.01
12 CM	1,1-Dichloroethene	0.494	0.467	5.5#	90	0.00
13 T	Acrolein	0.084	0.074	11.9	86	0.00
14 T	Allyl chloride	0.913	0.912	0.1	92	0.00
15 T	Acrylonitrile	0.299	0.318	-6.4	99	0.00
16 T	Acetone	0.263	0.275	-4.6	101	0.00
17 T	Carbon Disulfide	1.289	1.109	14.0	83	0.00
18 T	Methyl Acetate	0.770	0.878	-14.0	104	0.00
19 T	Methyl tert-butyl Ether	1.666	1.652	0.8	92	0.00
20 T	Methylene Chloride	0.563	0.536	4.8	93	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.500	8.1	88	0.00
22 T	Diisopropyl ether	1.752	1.744	0.5	93	0.00
23 T	Vinyl Acetate	1.263	1.172	7.2	91	0.00
24 P	1,1-Dichloroethane	0.961	0.953	0.8	94	0.00
25 T	2-Butanone	0.403	0.452	-12.2	103	0.00
26 T	2,2-Dichloropropane	0.774	0.648	16.3	77	0.00
27 T	cis-1,2-Dichloroethene	0.620	0.615	0.8	93	0.00
28 T	Bromochloromethane	0.427	0.480	-12.4	97	0.00
29 T	Tetrahydrofuran	0.263	0.293	-11.4	102	0.00
30 C	Chloroform	1.009	1.007	0.2#	93	0.00
31 T	Cyclohexane	0.842	0.836	0.7	89	0.00
32 T	1,1,1-Trichloroethane	0.872	0.873	-0.1	90	0.00
33 S	1,2-Dichloroethane-d4	0.670	0.626	6.6	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.422	0.423	-0.2	93	0.00
36 T	1,1-Dichloropropene	0.576	0.546	5.2	88	0.00
37 T	Ethyl Acetate	0.600	0.636	-6.0	97	0.00
38 T	Carbon Tetrachloride	0.603	0.596	1.2	89	0.00
39 T	Methylcyclohexane	0.673	0.666	1.0	88	0.00
40 TM	Benzene	1.696	1.726	-1.8	93	0.00
41 T	Methacrylonitrile	0.347	0.371	-6.9	100	0.00
42 TM	1,2-Dichloroethane	0.613	0.587	4.2	89	0.00
43 T	Isopropyl Acetate	0.939	0.979	-4.3	93	0.00
44 TM	Trichloroethene	0.526	0.526	0.0	92	0.00
45 C	1,2-Dichloropropane	0.416	0.432	-3.8#	95	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX051718\
 Data File : VX001803.D
 Acq On : 17 May 2018 20:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:59:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.292	0.293	-0.3	93	0.00
47 T	Bromodichloromethane	0.542	0.568	-4.8	92	0.00
48 T	Methyl methacrylate	0.494	0.526	-6.5	94	0.00
49 T	1,4-Dioxane	0.010	0.012	-20.0	116	0.00
50 S	Toluene-d8	1.511	1.493	1.2	87	0.00
51 T	4-Methyl-2-Pentanone	0.595	0.664	-11.6	98	0.00
52 CM	Toluene	1.090	1.103	-1.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.638	0.660	-3.4	89	0.00
54 T	cis-1,3-Dichloropropene	0.562	0.518	7.8	82	0.00
55 T	1,1,2-Trichloroethane	0.417	0.403	3.4	89	0.00
56 T	Ethyl methacrylate	0.591	0.579	2.0	85	0.00
57 T	1,3-Dichloropropane	0.697	0.653	6.3	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.339	-11.5	93	0.00
59 T	2-Hexanone	0.426	0.449	-5.4	92	0.00
60 T	Dibromochloromethane	0.427	0.433	-1.4	88	0.00
61 T	1,2-Dibromoethane	0.436	0.416	4.6	86	0.00
62 S	4-Bromofluorobenzene	0.544	0.496	8.8	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.651	0.669	-2.8	86	0.00
65 PM	Chlorobenzene	1.375	1.376	-0.1	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.486	0.518	-6.6	87	0.00
67 C	Ethyl Benzene	2.198	2.247	-2.2#	84	0.00
68 T	m/p-Xylenes	0.867	0.905	-4.4	85	0.00
69 T	o-Xylene	0.831	0.878	-5.7	85	0.00
70 T	Styrene	1.370	1.463	-6.8	85	0.00
71 P	Bromoform	0.396	0.437	-10.4	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.293	3.576	-8.6	85	0.00
74 T	N-amyl acetate	1.579	1.847	-17.0	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.988	1.065	-7.8	90	0.00
76 T	1,2,3-Trichloropropane	0.951	1.029	-8.2	85	0.00
77 T	Bromobenzene	0.968	1.018	-5.2	85	0.00
78 T	n-propylbenzene	3.710	3.931	-6.0	84	0.00
79 T	2-Chlorotoluene	2.219	2.347	-5.8	85	0.00
80 T	1,3,5-Trimethylbenzene	2.701	2.969	-9.9	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.272	0.296	-8.8	82	0.00
82 T	4-Chlorotoluene	2.614	2.697	-3.2	83	0.00
83 T	tert-Butylbenzene	2.688	2.924	-8.8	86	0.00
84 T	1,2,4-Trimethylbenzene	2.797	3.073	-9.9	85	0.00
85 T	sec-Butylbenzene	3.318	3.593	-8.3	85	0.00
86 T	p-Isopropyltoluene	3.071	3.281	-6.8	84	0.00
87 T	1,3-Dichlorobenzene	1.789	1.814	-1.4	84	0.00
88 T	1,4-Dichlorobenzene	1.850	1.814	1.9	83	0.00
89 T	n-Butylbenzene	2.656	2.645	0.4	79	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX051718\
Data File : VX001803.D
Acq On : 17 May 2018 20:54
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 18 04:59:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.477	0.532	-11.5	86	0.00
91 T	1,2-Dichlorobenzene	1.783	1.868	-4.8	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.211	0.247	-17.1	90	0.00
93 T	1,2,4-Trichlorobenzene	1.421	1.389	2.3	79	0.00
94 T	Hexachlorobutadiene	0.763	0.748	2.0	81	0.00
95 T	Naphthalene	3.694	4.072	-10.2	85	0.00
96 T	1,2,3-Trichlorobenzene	1.431	1.448	-1.2	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6