


Data Path : W:\HPCHEM1\MSVOA X\Data\VX051718\  
Data File : VX001781.D  
Acq On : 17 May 2018 09:33  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X051518W.M  
Title : SW846 8260  
Last Update : Tue May 15 01:36:22 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 17.6      | 3650    | PASS             |
| 75          | 95           | 30           | 60           | 47.7      | 9900    | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 20739   | PASS             |
| 96          | 95           | 5            | 9            | 6.6       | 1364    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.2       | 238     | PASS             |
| 174         | 95           | 50           | 100          | 96.6      | 20024   | PASS             |
| 175         | 174          | 5            | 9            | 7.0       | 1411    | PASS             |
| 176         | 174          | 95           | 101          | 97.1      | 19438   | PASS             |
| 177         | 176          | 5            | 9            | 7.3       | 1427    | PASS             |