

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001825.D
 Acq On : 18 May 2018 14:54
 Operator : JC/MD
 Sample : J2853-12DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW305-180507DL

Quant Time: May 19 04:52:35 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	343624	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	254444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	146679	41.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.72%	
35) Dibromofluoromethane	5.51	113	130506	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
50) Toluene-d8	8.72	98	465243	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
62) 4-Bromofluorobenzene	11.14	95	117793	31.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.00%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	155304	47.95	ug/l	90
44) Trichloroethene	7.22	130	6125	1.69	ug/l	98
64) Tetrachloroethene	9.34	164	37584	11.35	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001825.D
Acq On : 18 May 2018 14:54
Operator : JC/MD
Sample : J2853-12DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW305-180507DL

Quant Time: May 19 04:52:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

