

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001827.D
 Acq On : 18 May 2018 15:47
 Operator : JC/MD
 Sample : J2853-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY067MW013-180507

Quant Time: May 19 03:47:15 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	293321	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	413674	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	180392	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	218031	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
35) Dibromofluoromethane	5.51	113	186720	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
50) Toluene-d8	8.72	98	667884	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
62) 4-Bromofluorobenzene	11.14	95	192682	42.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.60%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	31770	8.74	ug/l	92
44) Trichloroethene	7.22	130	3476	0.80	ug/l	98
64) Tetrachloroethene	9.34	164	18257	4.22	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001827.D
Acq On : 18 May 2018 15:47
Operator : JC/MD
Sample : J2853-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY067MW013-180507

Quant Time: May 19 03:47:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

