

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001830.D
 Acq On : 18 May 2018 17:07
 Operator : JC/MD
 Sample : J2853-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3TB07-180507

Quant Time: May 21 10:01:42 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	433284	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	343065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	185208	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	230216	56.20	ug/l	0.00
Spiked Amount			Recovery	=	112.40%	
35) Dibromofluoromethane	5.51	113	197141	53.92	ug/l	0.00
Spiked Amount			Recovery	=	107.84%	
50) Toluene-d8	8.72	98	709213	54.16	ug/l	0.00
Spiked Amount			Recovery	=	108.32%	
62) 4-Bromofluorobenzene	11.14	95	202488	42.94	ug/l	0.00
Spiked Amount			Recovery	=	85.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001830.D
Acq On : 18 May 2018 17:07
Operator : JC/MD
Sample : J2853-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3TB07-180507

Quant Time: May 21 10:01:42 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

