

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001832.D
 Acq On : 18 May 2018 18:01
 Operator : JC/MD
 Sample : J2853-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW307-180507

Quant Time: May 19 04:21:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	324306	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	453850	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	359395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	196707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	235399	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
35) Dibromofluoromethane	5.51	113	203482	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
50) Toluene-d8	8.72	98	730061	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
62) 4-Bromofluorobenzene	11.14	95	209483	42.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.82%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	80396	20.01	ug/l	90
30) Chloroform	5.22	83	1580	0.24	ug/l	93
44) Trichloroethene	7.22	130	16879	3.53	ug/l	89
64) Tetrachloroethene	9.34	164	8090	1.73	ug/l	96
65) Chlorobenzene	10.15	112	2182	0.22	ug/l	95
91) 1,2-Dichlorobenzene	12.40	146	5617	0.80	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001832.D
Acq On : 18 May 2018 18:01
Operator : JC/MD
Sample : J2853-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW307-180507

Quant Time: May 19 04:21:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

