

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001866.D
 Acq On : 19 May 2018 10:00
 Operator : JC/MD
 Sample : PB109413ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109413ZHE#17

Quant Time: May 21 04:29:46 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	356476	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	151780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150613	42.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.18%	
35) Dibromofluoromethane	5.51	113	134578	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
50) Toluene-d8	8.72	98	499513	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
62) 4-Bromofluorobenzene	11.14	95	138650	35.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.48%	
Target Compounds						
16) Acetone	2.45	43	8055	5.73	ug/l	97
18) Methyl Acetate	2.78	43	28342	6.90	ug/l	100
20) Methylene Chloride	2.85	84	5228	1.74	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001866.D
Acq On : 19 May 2018 10:00
Operator : JC/MD
Sample : PB109413ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109413ZHE#17

Quant Time: May 21 04:29:46 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

