

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
 Data File : VX001869.D
 Acq On : 19 May 2018 11:20
 Operator : JC/MD
 Sample : PB109413ZHE#20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109413ZHE#20

Quant Time: May 21 04:35:19 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	275711	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	361455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	263636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	115368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155710	42.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.26%	
35) Dibromofluoromethane	5.51	113	138437	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
50) Toluene-d8	8.72	98	505283	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
62) 4-Bromofluorobenzene	11.14	95	119855	30.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.94%	
Target Compounds						
16) Acetone	2.45	43	8333	5.74	ug/l	91
18) Methyl Acetate	2.78	43	28026	6.60	ug/l	98
20) Methylene Chloride	2.86	84	4789	1.54	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051818\
Data File : VX001869.D
Acq On : 19 May 2018 11:20
Operator : JC/MD
Sample : PB109413ZHE#20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109413ZHE#20

Quant Time: May 21 04:35:19 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

