

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
 Data File : VX001873.D
 Acq On : 21 May 2018 13:17
 Operator : JC/MD
 Sample : J2853-02RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY074MW067-180507RE

Quant Time: May 22 02:01:27 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	290168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	393287	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	294715	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	159293	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183068	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
35) Dibromofluoromethane	5.51	113	162249	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	8.72	98	584825	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.14	95	159318	37.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.44%	
Target Compounds						
4) Vinyl Chloride	1.40	62	7775	2.286	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	6124	1.703	ug/l	75
40) Benzene	6.15	78	9198	0.690	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
Data File : VX001873.D
Acq On : 21 May 2018 13:17
Operator : JC/MD
Sample : J2853-02RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY074MW067-180507RE

Quant Time: May 22 02:01:27 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

