

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
 Data File : VX001875.D
 Acq On : 21 May 2018 14:10
 Operator : JC/MD
 Sample : J2884-07RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0457-180509RE

Quant Time: May 22 02:07:52 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	279642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	368840	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	274854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	138651	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182444	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
35) Dibromofluoromethane	5.51	113	163449	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
50) Toluene-d8	8.72	98	567787	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.14	95	153359	38.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.42%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	3745	1.081	ug/l	44
44) Trichloroethene	7.22	130	5096	1.312	ug/l	99
64) Tetrachloroethene	9.34	164	24593	6.873	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
Data File : VX001875.D
Acq On : 21 May 2018 14:10
Operator : JC/MD
Sample : J2884-07RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0457-180509RE

Quant Time: May 22 02:07:52 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

