

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
 Data File : VX001880.D
 Acq On : 21 May 2018 16:25
 Operator : JC/MD
 Sample : J2986-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0458-180514

Quant Time: May 22 02:25:25 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358792	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	259311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	127077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	179525	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
35) Dibromofluoromethane	5.51	113	160928	53.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.30%	
50) Toluene-d8	8.72	98	471893	43.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.04%	
62) 4-Bromofluorobenzene	11.14	95	144653	37.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.10%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3399	1.091	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	18250	5.540	ug/l	99
44) Trichloroethene	7.22	130	196647	52.064	ug/l	100
64) Tetrachloroethene	9.34	164	5175	1.533	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
Data File : VX001880.D
Acq On : 21 May 2018 16:25
Operator : JC/MD
Sample : J2986-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0458-180514

Quant Time: May 22 02:25:25 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

