

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
 Data File : VX001882.D
 Acq On : 21 May 2018 17:19
 Operator : JC/MD
 Sample : J2988-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3TB11-180515

Quant Time: May 22 03:14:11 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254295	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	341878	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	255654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	126661	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	179357	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	5.51	113	160729	55.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.42%	
50) Toluene-d8	8.73	98	484200	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
62) 4-Bromofluorobenzene	11.14	95	148239	39.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.68%	
Target Compounds						
4) Vinyl Chloride	1.40	62	10452	3.507	ug/l	97
16) Acetone	2.45	43	3884	2.901	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	34272	10.878	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052118\
Data File : VX001882.D
Acq On : 21 May 2018 17:19
Operator : JC/MD
Sample : J2988-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3TB11-180515

Quant Time: May 22 03:14:11 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

