

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
 Data File : VX009753.D
 Acq On : 21 May 2019 17:21
 Operator : JC/SP
 Sample : K2910-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0226

Quant Time: May 22 08:02:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	173701	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	277489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106481	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116305	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
35) Dibromofluoromethane	5.50	113	84779	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	8.71	98	354682	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.13	95	115991	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
Target Compounds						
16) Acetone	2.45	43	4900	3.848	ug/l	91
27) cis-1,2-Dichloroethene	4.60	96	7381	3.499	ug/l	91
40) Benzene	6.14	78	3632	0.455	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052119\
Data File : VX009753.D
Acq On : 21 May 2019 17:21
Operator : JC/SP
Sample : K2910-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-0226

Quant Time: May 22 08:02:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

