

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
 Data File : VX009757.D
 Acq On : 21 May 2019 18:54
 Operator : JC/SP
 Sample : K2973-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-TB-2019-1

Quant Time: May 22 08:10:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	272263	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	98347	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115593	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
35) Dibromofluoromethane	5.50	113	84076	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
50) Toluene-d8	8.71	98	346792	50.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.13	95	112879	44.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.92%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052119\
Data File : VX009757.D
Acq On : 21 May 2019 18:54
Operator : JC/SP
Sample : K2973-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-TB-2019-1

Quant Time: May 22 08:10:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

