

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
 Data File : VX009761.D
 Acq On : 21 May 2019 20:28
 Operator : JC/SP
 Sample : K2973-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-0158

Quant Time: May 22 08:19:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178731	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285712	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	105262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121072	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
35) Dibromofluoromethane	5.50	113	88323	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
50) Toluene-d8	8.71	98	361111	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.13	95	116537	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds						
16) Acetone	2.45	43	3755	2.866	ug/l	97
65) Chlorobenzene	10.13	112	1747	0.346	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052119\
Data File : VX009761.D
Acq On : 21 May 2019 20:28
Operator : JC/SP
Sample : K2973-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-0158

Quant Time: May 22 08:19:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

