

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
 Data File : VX009783.D
 Acq On : 22 May 2019 06:10
 Operator : JC/SP
 Sample : K2953-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17D_R2

Quant Time: May 22 08:58:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	173577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	278514	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249748	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108336	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117753	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	5.50	113	85721	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	8.71	98	352589	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	11.13	95	117667	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
Target Compounds						
16) Acetone	2.44	43	4625	3.635	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	7685	3.645	ug/l	89
44) Trichloroethene	7.21	130	1077	0.556	ug/l	98
64) Tetrachloroethene	9.34	164	1440	0.841	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052119\
Data File : VX009783.D
Acq On : 22 May 2019 06:10
Operator : JC/SP
Sample : K2953-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17D_R2

Quant Time: May 22 08:58:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

