

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
 Data File : VX009791.D
 Acq On : 22 May 2019 09:17
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 22 10:37:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.461	0.434	5.9	75	0.00
3 P	Chloromethane	0.629	0.664	-5.6	85	0.00
4 C	Vinyl Chloride	0.613	0.645	-5.2#	83	0.00
5 T	Bromomethane	0.368	0.373	-1.4	84	0.00
6 T	Chloroethane	0.394	0.414	-5.1	85	0.00
7 T	Trichlorofluoromethane	0.746	0.785	-5.2	84	0.00
8 T	Diethyl Ether	0.353	0.355	-0.6	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.453	3.2	77	0.00
10 T	Methyl Iodide	0.483	0.526	-8.9	80	0.00
11 T	Tert butyl alcohol	0.158	0.165	-4.4	83	0.00
12 CM	1,1-Dichloroethene	0.469	0.480	-2.3#	80	0.00
13 T	Acrolein	0.143	0.120	16.1	68	0.00
14 T	Allyl chloride	1.143	1.140	0.3	80	0.00
15 T	Acrylonitrile	0.386	0.401	-3.9	85	0.00
16 T	Acetone	0.367	0.353	3.8	79	0.00
17 T	Carbon Disulfide	1.317	1.407	-6.8	83	0.00
18 T	Methyl Acetate	0.872	0.861	1.3	82	0.00
19 T	Methyl tert-butyl Ether	1.822	1.875	-2.9	82	0.00
20 T	Methylene Chloride	0.593	0.594	-0.2	82	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.525	-3.1	82	0.00
22 T	Diisopropyl ether	2.199	2.328	-5.9	85	0.00
23 T	Vinyl Acetate	1.883	2.044	-8.6	84	0.00
24 P	1,1-Dichloroethane	1.085	1.114	-2.7	83	0.00
25 T	2-Butanone	0.570	0.596	-4.6	83	0.00
26 T	2,2-Dichloropropane	0.807	0.488	39.5#	48#	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.613	-1.0	82	0.00
28 T	Bromochloromethane	0.487	0.589	-20.9#	93	0.00
29 T	Tetrahydrofuran	0.366	0.399	-9.0	86	0.00
30 C	Chloroform	0.981	1.012	-3.2#	83	0.00
31 T	Cyclohexane	1.009	1.077	-6.7	84	0.00
32 T	1,1,1-Trichloroethane	0.817	0.838	-2.6	82	0.00
33 S	1,2-Dichloroethane-d4	0.683	0.716	-4.8	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.314	0.310	1.3	79	0.00
36 T	1,1-Dichloropropene	0.465	0.469	-0.9	84	0.00
37 T	Ethyl Acetate	0.647	0.677	-4.6	86	0.00
38 T	Carbon Tetrachloride	0.422	0.420	0.5	82	0.00
39 T	Methylcyclohexane	0.571	0.552	3.3	79	0.00
40 TM	Benzene	1.438	1.440	-0.1	83	0.00
41 T	Methacrylonitrile	0.385	0.388	-0.8	85	0.00
42 TM	1,2-Dichloroethane	0.516	0.528	-2.3	86	0.00
43 T	Isopropyl Acetate	1.036	1.065	-2.8	84	0.00
44 TM	Trichloroethene	0.348	0.334	4.0	81	0.00
45 C	1,2-Dichloropropane	0.407	0.407	0.0	82	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX052119\
 Data File : VX009791.D
 Acq On : 22 May 2019 09:17
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 22 10:37:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.236	0.232	1.7	83	0.00
47 T	Bromodichloromethane	0.472	0.475	-0.6	83	0.00
48 T	Methyl methacrylate	0.514	0.537	-4.5	85	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	85	0.00
50 S	Toluene-d8	1.269	1.248	1.7	77	0.00
51 T	4-Methyl-2-Pentanone	0.677	0.720	-6.4	87	0.00
52 CM	Toluene	0.874	0.876	-0.2#	83	0.00
53 T	t-1,3-Dichloropropene	0.533	0.503	5.6	75	0.00
54 T	cis-1,3-Dichloropropene	0.589	0.559	5.1	75	0.00
55 T	1,1,2-Trichloroethane	0.357	0.349	2.2	81	0.00
56 T	Ethyl methacrylate	0.637	0.649	-1.9	82	0.00
57 T	1,3-Dichloropropane	0.628	0.626	0.3	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.338	0.328	3.0	76	0.00
59 T	2-Hexanone	0.532	0.565	-6.2	86	0.00
60 T	Dibromochloromethane	0.348	0.353	-1.4	81	0.00
61 T	1,2-Dibromoethane	0.360	0.363	-0.8	83	0.00
62 S	4-Bromofluorobenzene	0.461	0.456	1.1	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.343	0.354	-3.2	87	0.00
65 PM	Chlorobenzene	1.000	0.977	2.3	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.359	1.1	82	0.00
67 C	Ethyl Benzene	1.841	1.840	0.1#	83	0.00
68 T	m/p-Xylenes	0.673	0.677	-0.6	83	0.00
69 T	o-Xylene	0.664	0.660	0.6	83	0.00
70 T	Styrene	1.154	1.157	-0.3	82	0.00
71 P	Bromoform	0.295	0.297	-0.7	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.636	3.516	3.3	82	0.00
74 T	N-amyl acetate	2.120	2.162	-2.0	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.355	1.309	3.4	82	0.00
76 T	1,2,3-Trichloropropane	1.160	1.130	2.6	83	0.00
77 T	Bromobenzene	0.902	0.879	2.5	84	0.00
78 T	n-propylbenzene	4.152	4.056	2.3	82	0.00
79 T	2-Chlorotoluene	2.562	2.423	5.4	83	0.00
80 T	1,3,5-Trimethylbenzene	3.061	2.993	2.2	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.453	0.383	15.5	69	0.00
82 T	4-Chlorotoluene	2.894	2.807	3.0	83	0.00
83 T	tert-Butylbenzene	2.978	2.818	5.4	81	0.00
84 T	1,2,4-Trimethylbenzene	3.073	3.030	1.4	83	0.00
85 T	sec-Butylbenzene	3.531	3.376	4.4	81	0.00
86 T	p-Isopropyltoluene	3.174	3.045	4.1	80	0.00
87 T	1,3-Dichlorobenzene	1.609	1.536	4.5	81	0.00
88 T	1,4-Dichlorobenzene	1.622	1.560	3.8	82	0.00
89 T	n-Butylbenzene	2.869	2.731	4.8	78	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052119\
Data File : VX009791.D
Acq On : 22 May 2019 09:17
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 22 10:37:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.541	0.531	1.8	81	0.00
91 T	1,2-Dichlorobenzene	1.639	1.574	4.0	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.299	1.6	80	0.00
93 T	1,2,4-Trichlorobenzene	1.121	1.062	5.3	78	0.00
94 T	Hexachlorobutadiene	0.572	0.530	7.3	78	0.00
95 T	Naphthalene	3.646	3.543	2.8	79	0.00
96 T	1,2,3-Trichlorobenzene	1.136	1.084	4.6	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5