


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX052119\
 Data File : VX009741.D
 Acq On : 21 May 2019 11:36
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Title : SW846 8260
 Last Update : Thu May 09 07:31:15 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	7665	PASS
75	95	30	60	52.2	18958	PASS
95	95	100	100	100.0	36293	PASS
96	95	5	9	7.3	2650	PASS
173	174	0.00	2	0.7	215	PASS
174	95	50	100	81.5	29565	PASS
175	174	5	9	7.9	2331	PASS
176	174	95	101	99.1	29290	PASS
177	176	5	9	6.3	1855	PASS