

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001892.D
 Acq On : 22 May 2018 14:35
 Operator : JC/MD
 Sample : VX0522WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0522WBL01

Quant Time: May 23 01:40:26 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	320574	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	153809	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165325	51.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.84%	
35) Dibromofluoromethane	5.51	113	131179	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
50) Toluene-d8	8.72	98	465795	48.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.14%	
62) 4-Bromofluorobenzene	11.14	95	148194	42.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001892.D
Acq On : 22 May 2018 14:35
Operator : JC/MD
Sample : VX0522WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0522WBL01

Quant Time: May 23 01:40:26 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

