

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001894.D
 Acq On : 22 May 2018 15:36
 Operator : JC/MD
 Sample : J2988-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3TB11-180515

Quant Time: May 24 11:41:42 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	325492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	464230	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	403137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	216932	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	256864	58.88	ug/l	0.00
Spiked Amount						
			Recovery	=	117.76%	
35) Dibromofluoromethane	5.51	113	206581	52.73	ug/l	0.00
Spiked Amount						
			Recovery	=	105.46%	
50) Toluene-d8	8.72	98	737764	52.58	ug/l	0.00
Spiked Amount						
			Recovery	=	105.16%	
62) 4-Bromofluorobenzene	11.14	95	236359	46.79	ug/l	0.00
Spiked Amount						
			Recovery	=	93.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001894.D
Acq On : 22 May 2018 15:36
Operator : JC/MD
Sample : J2988-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3TB11-180515

Quant Time: May 24 11:41:42 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

