

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001909.D
 Acq On : 22 May 2018 22:18
 Operator : JC/MD
 Sample : PB109525ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109525ZHE#16

Quant Time: May 23 02:31:00 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	236012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	320630	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	262749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154639	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
35) Dibromofluoromethane	5.51	113	127404	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	8.72	98	451932	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
62) 4-Bromofluorobenzene	11.14	95	126747	36.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.66%	
Target Compounds						
16) Acetone	2.45	43	9064	7.29	ug/l	100
18) Methyl Acetate	2.78	43	24884	6.85	ug/l	100
20) Methylene Chloride	2.86	84	6832	2.57	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001909.D
Acq On : 22 May 2018 22:18
Operator : JC/MD
Sample : PB109525ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109525ZHE#16

Quant Time: May 23 02:31:00 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

