

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001918.D
 Acq On : 23 May 2018 02:45
 Operator : JC/MD
 Sample : PB109456TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109456TB

Quant Time: May 23 09:00:29 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329565	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	268109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	141259	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151597	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
35) Dibromofluoromethane	5.51	113	125849	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
50) Toluene-d8	8.72	98	459342	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
62) 4-Bromofluorobenzene	11.15	95	133894	37.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.66%	
Target Compounds						
16) Acetone	2.45	43	9397	7.24	ug/l	98
18) Methyl Acetate	2.78	43	22770	6.00	ug/l	100
20) Methylene Chloride	2.86	84	6140	2.21	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001918.D
Acq On : 23 May 2018 02:45
Operator : JC/MD
Sample : PB109456TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109456TB

Quant Time: May 23 09:00:29 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

