

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001921.D
 Acq On : 23 May 2018 04:06
 Operator : JC/MD
 Sample : PB109456ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109456ZHE#01

Quant Time: May 23 05:02:58 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	346640	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	296098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	173115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161533	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
35) Dibromofluoromethane	5.51	113	134148	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
50) Toluene-d8	8.72	98	487324	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
62) 4-Bromofluorobenzene	11.14	95	153031	40.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.14%	
Target Compounds						
16) Acetone	2.45	43	9205	6.83	ug/l	99
18) Methyl Acetate	2.78	43	23061	5.85	ug/l	99
20) Methylene Chloride	2.86	84	5955	2.06	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001921.D
Acq On : 23 May 2018 04:06
Operator : JC/MD
Sample : PB109456ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109456ZHE#01

Quant Time: May 23 05:02:58 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

