

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001926.D
 Acq On : 23 May 2018 06:19
 Operator : JC/MD
 Sample : PB109456ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109456ZHE#06

Quant Time: May 23 06:57:35 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	336484	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	278712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153724	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	156828	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
35) Dibromofluoromethane	5.51	113	130586	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	8.72	98	478019	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
62) 4-Bromofluorobenzene	11.14	95	139452	38.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.16%	
Target Compounds						
16) Acetone	2.45	43	8895	6.79	ug/l	96
18) Methyl Acetate	2.78	43	23335	6.10	ug/l	99
20) Methylene Chloride	2.86	84	6389	2.28	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001926.D
Acq On : 23 May 2018 06:19
Operator : JC/MD
Sample : PB109456ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109456ZHE#06

Quant Time: May 23 06:57:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

