

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
 Data File : VX001927.D
 Acq On : 23 May 2018 06:46
 Operator : JC/MD
 Sample : PB109456ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109456ZHE#10

Quant Time: May 23 11:51:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	344187	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	159891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159202	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
35) Dibromofluoromethane	5.51	113	132665	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
50) Toluene-d8	8.72	98	481372	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
62) 4-Bromofluorobenzene	11.14	95	144328	38.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.06%	
Target Compounds						
16) Acetone	2.45	43	9779	7.40	ug/l	97
18) Methyl Acetate	2.78	43	22464	5.82	ug/l	98
20) Methylene Chloride	2.86	84	7439	2.63	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052218\
Data File : VX001927.D
Acq On : 23 May 2018 06:46
Operator : JC/MD
Sample : PB109456ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109456ZHE#10

Quant Time: May 23 11:51:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

