

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052418\
 Data File : VX001942.D
 Acq On : 24 May 2018 15:40
 Operator : JC/MD
 Sample : VX0524WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0524WBL01

Quant Time: May 25 03:44:58 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	381541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	543745	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	490731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	264131	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	271021	39.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.48%	
35) Dibromofluoromethane	5.51	113	212070	39.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.00%	
50) Toluene-d8	8.72	98	774263	39.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.66%	
62) 4-Bromofluorobenzene	11.14	95	286523	37.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052418\
Data File : VX001942.D
Acq On : 24 May 2018 15:40
Operator : JC/MD
Sample : VX0524WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0524WBL01

Quant Time: May 25 03:44:58 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

