

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052418\
 Data File : VX001951.D
 Acq On : 24 May 2018 19:50
 Operator : JC/MD
 Sample : J3060-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-WC-1

Quant Time: May 25 04:11:51 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	346690	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	530494	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	484649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	265285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	267507	43.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.34%	
35) Dibromofluoromethane	5.51	113	208745	39.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.70%	
50) Toluene-d8	8.72	98	758114	39.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.94%	
62) 4-Bromofluorobenzene	11.14	95	281912	37.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.04%	
Target Compounds						
16) Acetone	2.45	43	15531	6.49	ug/l	98
18) Methyl Acetate	2.79	43	30208	4.20	ug/l	100
20) Methylene Chloride	2.86	84	69642	15.67	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	24191	1.22	ug/l	100
95) Naphthalene	13.84	128	230909	9.64	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052418\
Data File : VX001951.D
Acq On : 24 May 2018 19:50
Operator : JC/MD
Sample : J3060-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-WC-1

Quant Time: May 25 04:11:51 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

