

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX001973.D
 Acq On : 25 May 2018 13:30
 Operator : JC/MD
 Sample : J3076-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-20180522

Quant Time: May 26 01:51:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	327017	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	491234	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	444850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	234377	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	295490	50.55	ug/l	0.00
Spiked Amount			Recovery	=	101.10%	
35) Dibromofluoromethane	5.51	113	229318	46.68	ug/l	0.00
Spiked Amount			Recovery	=	93.36%	
50) Toluene-d8	8.72	98	838244	47.13	ug/l	0.00
Spiked Amount			Recovery	=	94.26%	
62) 4-Bromofluorobenzene	11.14	95	303849	43.67	ug/l	0.00
Spiked Amount			Recovery	=	87.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX001973.D
Acq On : 25 May 2018 13:30
Operator : JC/MD
Sample : J3076-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-20180522

Quant Time: May 26 01:51:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

