

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX001978.D
 Acq On : 25 May 2018 15:43
 Operator : JC/MD
 Sample : J3094-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HS-BASELINE-WATER

Quant Time: May 26 02:14:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	306919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	472153	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	415767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	216402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	293359	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
35) Dibromofluoromethane	5.51	113	225752	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
50) Toluene-d8	8.72	98	790698	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
62) 4-Bromofluorobenzene	11.14	95	283509	42.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.78%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	15691	7.411	ug/l	97
29) Tetrahydrofuran	5.18	42	28291	11.205	ug/l	99
51) 4-Methyl-2-Pentanone	8.66	43	122565	14.252	ug/l	100
52) Toluene	8.79	92	20313	1.659	ug/l	99
68) m/p-Xylenes	10.36	106	5065	0.579	ug/l	92
69) o-Xylene	10.70	106	14216	1.627	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	4351	0.269	ug/l	85
95) Naphthalene	13.84	128	16026	0.820	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX001978.D
Acq On : 25 May 2018 15:43
Operator : JC/MD
Sample : J3094-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HS-BASELINE-WATER

Quant Time: May 26 02:14:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

