

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX001980.D
 Acq On : 25 May 2018 16:36
 Operator : JC/MD
 Sample : J3015-15
 Misc : 5.48g/5mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC7-BTM-5-5.5

Manual Integrations
 APPROVED

MMDadoda
 5/29/2018 4:39:23 PM

Quant Time: May 26 03:19:33 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	308412	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	465728	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	453745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	279799	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	278486	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
35) Dibromofluoromethane	5.51	113	215765	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	8.72	98	814673	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
62) 4-Bromofluorobenzene	11.15	95	326317	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.46	83	7132	0.896	ug/l	92
67) Ethyl Benzene	10.26	91	32010	1.282	ug/l	94
68) m/p-Xylenes	10.37	106	24755	2.593	ug/l	97
69) o-Xylene	10.71	106	32519	3.411	ug/l	98
73) Isopropylbenzene	11.02	105	37006	1.562	ug/l	99
78) n-propylbenzene	11.37	91	120472	4.540	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	268843	13.277	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	990104	47.318	ug/l	98
85) sec-Butylbenzene	11.95	105	44122	1.847	ug/l	98
86) p-Isopropyltoluene	12.07	119	35790	1.675	ug/l	96
89) n-Butylbenzene	12.39	91	181935m	10.046	ug/l	
95) Naphthalene	13.84	128	356056	14.090	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX001980.D
Acq On : 25 May 2018 16:36
Operator : JC/MD
Sample : J3015-15
Misc : 5.48g/5mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC7-BTM-5-5.5

Manual Integrations
APPROVED

MMDadoda
5/29/2018 4:39:23 PM

Quant Time: May 26 03:19:33 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

