

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX001985.D
 Acq On : 25 May 2018 18:50
 Operator : JC/MD
 Sample : J3064-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: May 26 03:39:37 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	314474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	477206	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	436383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	240641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289755	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
35) Dibromofluoromethane	5.51	113	225213	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
50) Toluene-d8	8.72	98	818728	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
62) 4-Bromofluorobenzene	11.14	95	301874	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
Target Compounds						
16) Acetone	2.45	43	13475	6.211	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX001985.D
Acq On : 25 May 2018 18:50
Operator : JC/MD
Sample : J3064-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: May 26 03:39:37 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

