

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX001986.D
 Acq On : 25 May 2018 19:16
 Operator : JC/MD
 Sample : J3064-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16

Manual Integrations
 APPROVED

MMDadoda
 5/29/2018 4:39:23 PM

Quant Time: May 26 03:43:39 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	319242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	481247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	441598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	249514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	291995	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	5.51	113	225426	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
50) Toluene-d8	8.72	98	818423	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
62) 4-Bromofluorobenzene	11.14	95	306172	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.08	59	14204	12.681	ug/l	98
16) Acetone	2.45	43	25281	11.479	ug/l	100
19) Methyl tert-butyl Ether	3.21	73	19783	1.414	ug/l #	90
39) Methylcyclohexane	7.47	83	5099	0.620	ug/l #	93
73) Isopropylbenzene	11.02	105	102610	4.856	ug/l	98
78) n-propylbenzene	11.37	91	142513	6.022	ug/l	98
83) tert-Butylbenzene	11.77	119	22341	1.261	ug/l	91
85) sec-Butylbenzene	11.95	105	116081	5.450	ug/l	82
89) n-Butylbenzene	12.39	91	31356m	1.942	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX001986.D
Acq On : 25 May 2018 19:16
Operator : JC/MD
Sample : J3064-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16

Manual Integrations
APPROVED

MMDadoda
5/29/2018 4:39:23 PM

Quant Time: May 26 03:43:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

