

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
 Data File : VX002003.D
 Acq On : 26 May 2018 03:42
 Operator : JC/MD
 Sample : PB109661ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109661ZHE#11

Quant Time: May 26 04:51:12 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	293808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	441246	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	406799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	218013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	265897	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
35) Dibromofluoromethane	5.51	113	205568	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	8.72	98	750574	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
62) 4-Bromofluorobenzene	11.14	95	275316	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
16) Acetone	2.45	43	15207	7.503	ug/l	98
18) Methyl Acetate	2.79	43	31236	5.383	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052518\
Data File : VX002003.D
Acq On : 26 May 2018 03:42
Operator : JC/MD
Sample : PB109661ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109661ZHE#11

Quant Time: May 26 04:51:12 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

