

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
 Data File : VX002025.D
 Acq On : 26 May 2018 18:06
 Operator : JC/MD
 Sample : J3118-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0460-180521

Quant Time: May 29 04:35:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	309987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	432913	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	379376	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	199962	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	234438	42.31	ug/l	0.00
Spiked Amount			Recovery	=	84.62%	
35) Dibromofluoromethane	5.51	113	185896	42.94	ug/l	0.00
Spiked Amount			Recovery	=	85.88%	
50) Toluene-d8	8.72	98	704572	44.95	ug/l	0.00
Spiked Amount			Recovery	=	89.90%	
62) 4-Bromofluorobenzene	11.14	95	245017	39.96	ug/l	0.00
Spiked Amount			Recovery	=	79.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	1551	0.34	ug/l #	87
12) 1,1-Dichloroethene	2.38	96	26175	7.24	ug/l	96
24) 1,1-Dichloroethane	3.70	63	3609	0.49	ug/l #	85
27) cis-1,2-Dichloroethene	4.61	96	52227	11.08	ug/l	99
42) 1,2-Dichloroethane	6.21	62	2118	0.30	ug/l #	76
44) Trichloroethene	7.21	130	20806	4.05	ug/l	96
64) Tetrachloroethene	9.34	164	56448	10.38	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
Data File : VX002025.D
Acq On : 26 May 2018 18:06
Operator : JC/MD
Sample : J3118-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0460-180521

Quant Time: May 29 04:35:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

