

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
 Data File : VX002030.D
 Acq On : 26 May 2018 20:19
 Operator : JC/MD
 Sample : J3118-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052OW376-180522

Quant Time: May 28 03:08:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	314157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	431081	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	382722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	205071	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	262224	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
35) Dibromofluoromethane	5.51	113	205417	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
50) Toluene-d8	8.72	98	777223	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.14	95	271347	44.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.88%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1202	0.26	ug/l	# 82
19) Methyl tert-butyl Ether	3.21	73	68002	4.94	ug/l	100
27) cis-1,2-Dichloroethene	4.61	96	35767	7.49	ug/l	97
44) Trichloroethene	7.22	130	16650	3.25	ug/l	89
64) Tetrachloroethene	9.34	164	8646	1.58	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
Data File : VX002030.D
Acq On : 26 May 2018 20:19
Operator : JC/MD
Sample : J3118-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052OW376-180522

Quant Time: May 28 03:08:31 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

