

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
 Data File : VX002033.D
 Acq On : 26 May 2018 21:39
 Operator : JC/MD
 Sample : J3118-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW732R-180521

Quant Time: May 29 05:14:07 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	310087	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	428824	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	377537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	203866	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	261698	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
35) Dibromofluoromethane	5.51	113	204238	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	8.72	98	770085	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	11.14	95	270808	44.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.16%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	3580	0.99	ug/l #	86
17) Carbon Disulfide	2.57	76	4445	0.43	ug/l #	93
27) cis-1,2-Dichloroethene	4.60	96	2496	0.53	ug/l #	1
40) Benzene	6.15	78	8022	0.46	ug/l #	83
64) Tetrachloroethene	9.34	164	2958	0.55	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
Data File : VX002033.D
Acq On : 26 May 2018 21:39
Operator : JC/MD
Sample : J3118-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW732R-180521

Quant Time: May 29 05:14:07 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

