

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
 Data File : VX002043.D
 Acq On : 27 May 2018 02:58
 Operator : JC/MD
 Sample : J3094-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HS-RO-7212003

Quant Time: May 29 05:47:29 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	13871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	27555	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	19327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	8141	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	60684	244.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	489.50%	
35) Dibromofluoromethane	5.51	113	40575	147.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	294.50%	
50) Toluene-d8	8.72	98	71052	71.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	142.44%	
62) 4-Bromofluorobenzene	11.14	95	22964	58.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.66%	
Target Compounds						
16) Acetone	2.45	43	17508	182.97	ug/l	97
18) Methyl Acetate	2.78	43	25275	111.03	ug/l	96
20) Methylene Chloride	2.86	84	19649	120.96	ug/l	96
30) Chloroform	5.23	83	5740	15.54	ug/l	97
95) Naphthalene	13.84	128	102316	139.16	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
Data File : VX002043.D
Acq On : 27 May 2018 02:58
Operator : JC/MD
Sample : J3094-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HS-RO-7212003

Quant Time: May 29 05:47:29 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

