

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
 Data File : VX002047.D
 Acq On : 27 May 2018 04:44
 Operator : JC/MD
 Sample : J3104-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-B

Quant Time: May 29 06:02:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	29738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	45426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	41365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	19933	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	64258	120.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	241.76%	
35) Dibromofluoromethane	5.51	113	47710	105.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	210.06%	
50) Toluene-d8	8.72	98	165078	100.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	200.76%	
62) 4-Bromofluorobenzene	11.14	95	56400	87.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	175.30%	

Target Compounds						Qvalue
16) Acetone	2.45	43	15499	75.55	ug/l	98
18) Methyl Acetate	2.78	43	27949	56.71	ug/l	97
20) Methylene Chloride	2.86	84	6623	17.56	ug/l	88
30) Chloroform	5.22	83	2117	2.67	ug/l	96
68) m/p-Xylenes	10.37	106	1531	1.76	ug/l	89
80) 1,3,5-Trimethylbenzene	11.51	105	2739	1.90	ug/l	93
84) 1,2,4-Trimethylbenzene	11.81	105	8829	5.92	ug/l	96
95) Naphthalene	13.84	128	22628	12.57	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052618\
Data File : VX002047.D
Acq On : 27 May 2018 04:44
Operator : JC/MD
Sample : J3104-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-B

Quant Time: May 29 06:02:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

