

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002065.D
 Acq On : 27 May 2018 13:46
 Operator : JC/MD
 Sample : J3094-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HS-RO-7212003

Quant Time: May 29 03:33:05 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	448817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	401854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	219400	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	210250	36.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.96%	
35) Dibromofluoromethane	5.51	113	164011	36.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.08%	
50) Toluene-d8	8.72	98	639083	39.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.66%	
62) 4-Bromofluorobenzene	11.14	95	225329	35.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.88%	
Target Compounds						
16) Acetone	2.45	43	16345	7.35	ug/l	93
18) Methyl Acetate	2.78	43	22081	3.05	ug/l	93
20) Methylene Chloride	2.86	84	52594	12.40	ug/l	98
95) Naphthalene	13.84	128	128968	6.51	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002065.D
Acq On : 27 May 2018 13:46
Operator : JC/MD
Sample : J3094-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HS-RO-7212003

Quant Time: May 29 03:33:05 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

