

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002071.D
 Acq On : 27 May 2018 16:25
 Operator : JC/MD
 Sample : J3099-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 200027

Quant Time: May 29 03:55:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	320445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	454749	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	412675	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	236568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213269	37.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.46%	
35) Dibromofluoromethane	5.51	113	166046	36.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.02%	
50) Toluene-d8	8.72	98	642613	39.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.06%	
62) 4-Bromofluorobenzene	11.14	95	235554	36.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.14%	
Target Compounds						
16) Acetone	2.45	43	18326	8.29	ug/l	97
18) Methyl Acetate	2.78	43	26890	4.00	ug/l	99
20) Methylene Chloride	2.86	84	6580497	1776.84	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002071.D
Acq On : 27 May 2018 16:25
Operator : JC/MD
Sample : J3099-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
200027

Quant Time: May 29 03:55:44 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

