

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002078.D
 Acq On : 27 May 2018 19:31
 Operator : JC/MD
 Sample : J3130-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5-WC-3

Quant Time: May 29 08:37:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	318574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	444370	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	402916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	227171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	208296	36.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.16%	
35) Dibromofluoromethane	5.51	113	162589	36.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.18%	
50) Toluene-d8	8.72	98	634935	39.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.94%	
62) 4-Bromofluorobenzene	11.14	95	228485	36.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.60%	
Target Compounds						
8) Diethyl Ether	2.19	74	8910	3.46	ug/l	Qvalue 100
16) Acetone	2.45	43	19044	8.67	ug/l	98
18) Methyl Acetate	2.78	43	25198	3.71	ug/l	97
20) Methylene Chloride	2.86	84	64445	15.80	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	23927	1.41	ug/l	98
95) Naphthalene	13.84	128	531383	25.90	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002078.D
Acq On : 27 May 2018 19:31
Operator : JC/MD
Sample : J3130-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5-WC-3

Quant Time: May 29 08:37:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

