

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002079.D
 Acq On : 27 May 2018 19:58
 Operator : JC/MD
 Sample : J3130-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5-WC-4

Quant Time: May 29 04:29:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	321078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	455932	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	408648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	227246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	214003	37.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.58%	
35) Dibromofluoromethane	5.51	113	164769	36.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.28%	
50) Toluene-d8	8.72	98	641113	38.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.68%	
62) 4-Bromofluorobenzene	11.14	95	231497	35.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.70%	
Target Compounds						
16) Acetone	2.45	43	15826	7.15	ug/l	97
18) Methyl Acetate	2.78	43	28787	4.36	ug/l	97
20) Methylene Chloride	2.86	84	128783	33.01	ug/l	98
25) 2-Butanone	4.71	43	5108	1.26	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	20526	1.21	ug/l	100
95) Naphthalene	13.84	128	227190	11.07	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002079.D
Acq On : 27 May 2018 19:58
Operator : JC/MD
Sample : J3130-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5-WC-4

Quant Time: May 29 04:29:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

