

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002082.D
 Acq On : 27 May 2018 21:17
 Operator : JC/MD
 Sample : J3114-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NYS-RT-3460

Manual Integrations
 APPROVED

MMDadoda
 5/30/2018 1:50:36 PM

Quant Time: May 29 08:52:40 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	329299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	463192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	431518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	248140	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	216787	36.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.66%	
35) Dibromofluoromethane	5.51	113	168973	36.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.96%	
50) Toluene-d8	8.72	98	666613	39.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.50%	
62) 4-Bromofluorobenzene	11.14	95	247081	37.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.32%	
Target Compounds						
16) Acetone	2.45	43	18557	8.17	ug/l	95
20) Methylene Chloride	2.82	84	48794150m	12831.65	ug/l	
28) Bromochloromethane	5.04	49	23071	5.22	ug/l	98
95) Naphthalene	13.84	128	98485	4.39	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002082.D
Acq On : 27 May 2018 21:17
Operator : JC/MD
Sample : J3114-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NYS-RT-3460

Manual Integrations
APPROVED

MMDadoda
5/30/2018 1:50:36 PM

Quant Time: May 29 08:52:40 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

