

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
 Data File : VX002083.D
 Acq On : 27 May 2018 21:44
 Operator : JC/MD
 Sample : J3114-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NYS-RBR-200035

Quant Time: May 29 08:56:17 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	440311	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	391323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	216881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207787	37.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.06%	
35) Dibromofluoromethane	5.51	113	161980	36.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.58%	
50) Toluene-d8	8.72	98	619271	38.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.70%	
62) 4-Bromofluorobenzene	11.14	95	221903	35.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.16%	
Target Compounds						
16) Acetone	2.45	43	23219	10.72	ug/l	98
18) Methyl Acetate	2.78	43	28459	4.42	ug/l	98
20) Methylene Chloride	2.86	84	252740	68.00	ug/l	98
95) Naphthalene	13.84	128	266993	13.63	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052718\
Data File : VX002083.D
Acq On : 27 May 2018 21:44
Operator : JC/MD
Sample : J3114-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NYS-RBR-200035

Quant Time: May 29 08:56:17 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

