

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052918\
 Data File : VX002094.D
 Acq On : 29 May 2018 15:46
 Operator : JC/MD
 Sample : J3118-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0460-180521

Quant Time: May 30 13:07:04 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	302448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	426723	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	380640	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208133	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	268561	49.68	ug/l	0.00
Spiked Amount			Recovery	=	99.36%	
35) Dibromofluoromethane	5.51	113	205207	48.09	ug/l	0.00
Spiked Amount			Recovery	=	96.18%	
50) Toluene-d8	8.72	98	787296	50.96	ug/l	0.00
Spiked Amount			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.14	95	279550	46.25	ug/l	0.00
Spiked Amount			Recovery	=	92.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	2287	0.51	ug/l	98
12) 1,1-Dichloroethene	2.38	96	43197	12.25	ug/l	96
24) 1,1-Dichloroethane	3.70	63	5473	0.77	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	82459	17.94	ug/l	90
42) 1,2-Dichloroethane	6.20	62	3083	0.44	ug/l	100
44) Trichloroethene	7.22	130	34378	6.78	ug/l	99
64) Tetrachloroethene	9.34	164	93356	17.11	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052918\
Data File : VX002094.D
Acq On : 29 May 2018 15:46
Operator : JC/MD
Sample : J3118-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0460-180521

Quant Time: May 30 13:07:04 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

