

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052918\
 Data File : VX002125.D
 Acq On : 30 May 2018 06:26
 Operator : JC/MD
 Sample : PB109723ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109723ZHE#06

Quant Time: May 30 11:31:28 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	294498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	425801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	402961	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	247257	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	205944	39.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.24%	
35) Dibromofluoromethane	5.51	113	149397	35.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.18%	
50) Toluene-d8	8.72	98	613536	39.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.60%	
62) 4-Bromofluorobenzene	11.14	95	234983	38.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	9048	4.45	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX052918\
Data File : VX002125.D
Acq On : 30 May 2018 06:26
Operator : JC/MD
Sample : PB109723ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109723ZHE#06

Quant Time: May 30 11:31:28 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

